

# Near-Infrared Spectroscopy: A Promising Prehospital Tool for Management of Traumatic Brain Injury

Q1 Joost Peters, MD<sup>1,2</sup> Bas Van Wageningen, MD;<sup>1,2</sup> Nico Hoogerwerf, MD, PhD;<sup>2,3</sup> Edward Tan, MD, PhD<sup>1,2</sup>

1. Department of Trauma Surgery, Radboud University Medical Center Nijmegen, Geert Grooteplein Zuid 10, 6525 GA The Netherlands
2. Helicopter Emergency Medical Service, Radboud University Medical Center Nijmegen, Geert Grooteplein Zuid 10, 6525 GA The Netherlands
3. Department of Anaesthesiology, Radboud University Medical Center Nijmegen, Geert Grooteplein Zuid 10, 6525 GA The Netherlands

## Correspondence:

Joost Peters, MD  
Department of Surgery  
Radboud University Medical Center  
Geert Grooteplein Zuid 10  
6525 GA Nijmegen, The Netherlands  
E-mail: Joost.Peters@radboudumc.nl

**Conflicts of interest/funding:** The Infrascanner used was provided by InfraScan Inc. (Philadelphia, Pennsylvania USA) free of charge. No funding or research support was received by the HEMS, the research team, or the authors of this article. A portion of this article was previously published in Dutch.<sup>18</sup> Permission to publish these data was granted by the publisher. None of the authors had competing interests in relation to this study.

**Keywords:** diagnostic equipment; infrared spectrophotometry; prehospital emergency care; traumatic brain injury

## Abbreviations:

CT: computed tomography  
EMS: Emergency Medical Service  
GCS: Glasgow Coma Score  
HEMS: helicopter Emergency Medical Service  
NIRS: near-infrared spectroscopy  
TBI: traumatic brain injury

## Abstract

**Introduction:** Early identification of traumatic brain injury (TBI) is essential. Near-infrared spectroscopy (NIRS) can be used in prehospital settings for non-invasive monitoring and the diagnosis of patients who may require surgical intervention.

**Methods:** The handheld NIRS Infrascanner (InfraScan Inc.; Philadelphia, Pennsylvania USA) uses eight symmetrical scan points to detect intracranial bleeding. A scanner was tested in a physician-staffed helicopter Emergency Medical Service (HEMS). The results were compared with those obtained using in-hospital computed tomography (CT) scans. Scan time, ease-of-use, and change in treatment were scored.

**Results:** A total of 25 patients were included. Complete scans were performed in 60% of patients. In 15 patients, the scan was abnormal, and in one patient, the scan resulted in a treatment change. Compared with the results of CT scanning, the Infrascanner obtained a sensitivity of 93.3% and a specificity of 78.6%. Most patients had severe TBI with indication for transport to a trauma center prior to scanning. In one patient, the scan resulted in a treatment change. Evaluation of patients with less severe TBI is needed to support the usefulness of the Infrascanner as a prehospital triage tool.

**Conclusion:** Promising results were obtained using the InfraScan NIRS device in prehospital screening for intracranial hematomas in TBI patients. High sensitivity and good specificity were found. Further research is necessary to determine the beneficial effects of enhanced prehospital screening on triage, survival, and quality of life in TBI patients.

Peters J, Van Wageningen B, Hoogerwerf N, Tan E. Near-infrared spectroscopy: a promising prehospital tool for management of traumatic brain injury. *Prehosp Disaster Med.* 2017;32(3):1-5.

## Introduction

Worldwide, trauma is one of the leading causes of death, disability, and health care costs.<sup>1,2</sup> In particular, trauma resulting in traumatic brain injury (TBI) is a major burden for society.<sup>1</sup> Early identification of TBI is essential. In addition to optimizing treatment strategies, immediate transport to an appropriate-level trauma center with neurosurgical intervention options improves the outcome for TBI patients.<sup>3-5</sup>

In patients with low Glasgow Coma Scores (GCS), the decision to transport to a Level 1 trauma center is evident. However, this decision is less clear when patients have a temporarily lower GCS, post-traumatic amnesia, or are intoxicated. Adequate patient selection is important, but difficult, when using only clinical parameters. For instance, a patient with an expanding epidural hematoma may initially present with no/mild complaints, but this condition can change rapidly. Prehospital on-site cerebral scanning for a hematoma could optimize patient triage, allow the initiation of brain-preserving resuscitation, and potentially improve the outcome for TBI patients.

Near-infrared spectroscopy (NIRS) is a technique that uses the reflection of light to detect fluid/hemoglobin presence near the brain. Near-infrared spectroscopy has been described

Received: May 24, 2016

Revised: September 13, 2016

Accepted: October 7, 2016

doi:10.1017/S1049023X17006367



Joost © 2017 Prehospital and Disaster Medicine

**Figure 1.** The Infrascanner Model 2000 (InfraScan Inc.; Philadelphia, Pennsylvania USA).

as a promising “work in progress” when used for non-invasive cerebral tissue monitoring.<sup>6</sup> Near-infrared is the term used to define light with a wavelength of approximately 600 to 1,000 nm. In this range, tissues are relatively translucent because of the low absorption of the waves. The amount of absorption of this energy is used to determine the tissue or fluid type because these light waves are able to penetrate several centimeters into tissues, including bone. Using this measurement, NIRS can be used for non-invasive monitoring and diagnosing. When comparing the reflection patterns on both sides of the patient’s skull, an asymmetrical reading can indicate a subdural, epidural, or extra-cranial haematoma.<sup>7</sup>

The purpose of this methodological clinical study was to evaluate the practical use of the NIRS Infrascanner device in the prehospital on-site screening of patients suspected of having a TBI. This screening was conducted by members of the physician-staffed helicopter Emergency Medical Service (HEMS).

#### Patients and Methods

The Infrascanner Model 2000 (InfraScan Inc.; Philadelphia, Pennsylvania USA) is a handheld NIRS apparatus (Figure 1). A precursor of this device was developed with the support of the United States Navy’s Office of Naval Research (Arlington, Virginia USA) and the US Marine Corps (Arlington, Virginia USA) and has Food and Drug Administration (Silver Spring,



Joost © 2017 Prehospital and Disaster Medicine

**Figure 2.** Scanning a Patient’s Head.

Maryland USA) clearance,<sup>8,9</sup> which was obtained after clinical testing and validation in a large, double-blinded, multi-center trial.<sup>10</sup> The Infrascanner was used in TBI patients prior to computed tomography (CT) scans of the brain, which represent the gold standard for detecting intracerebral hematomas. In 365 patients, a sensitivity of 88% and a specificity of 91% were demonstrated for hematomas larger than 3.5 cc in volume and less than 2.5 cm from the surface.<sup>10</sup> Other studies of smaller patient groups and of children confirmed the usefulness of the Infrascanner in detecting TBIs.<sup>11–14</sup>

To perform an Infrascanner examination, eight symmetrical points (bilateral frontal, temporal, parietal, and occipital) on the patient’s head must be scanned (Figure 2). The measured data are presented on a screen, and an abnormal reading in comparison with the contralateral side is presented in red.

After approval of the local ethical committee (Radboud University Medical Center; the Netherlands), a feasibility study was started with an Infrascanner 2000 in a regional HEMS in the southeast of the Netherlands. In the Netherlands, four HEMS function as an adjunct to paramedic ambulance services, according to dispatch guidelines. These HEMS operate 24/7 using night vision goggles. Dutch paramedic ambulance services (with registered Emergency Medical Service [EMS] nurses) work according to national EMS guidelines. These guidelines state that registered EMS nurses are not allowed to perform advanced prehospital treatment, such as rapid sequence induction using muscle relaxants, nor are they trained to perform thoracostomies for penetrating thoracic trauma. For these interventions in severely injured patients, the physician-staffed HEMS provides additional on-site knowledge and treatment options.

The HEMS physicians (n = 8) were adequately trained and familiarized with the scanner, prior to its operational use, by a salesman in cooperation with an author of the manuscript (ET).

Patients with suspected neurocranial trauma were scanned, including patients with normal and diminished GCS. All consecutive patients with TBI could be included, but frequently

a scan was not conducted because other treatment priorities. These patients were excluded from the research cohort. Hospital CT scan results were available and were compared with the InfraScan results. Data were collected using the dispatch database combined with a questionnaire regarding ease-of-use and satisfaction with the scanner. In this questionnaire, the HEMS physician scored the scan time, scan results, and ease-of-use on a zero to 10 scale. It was recorded if the scan results led to a change in treatment or hospital choice.

Trauma patient often have suspected cervical spine injuries. Because of the immobilization that is needed in these patients, access to the dorsal aspects of the head is limited. This might lead to incomplete scans but reflects the scanner's actual working environment. To evaluate the scanner's usability in prehospital setting, all data, including the partial complete scans, were analyzed. Data were collected and analyzed using Microsoft Office Excel 2007 (Microsoft Corporation; Redmond, Washington USA) and are presented as medians with ranges.

## Results

A total of 25 patients were included in the study, most of whom were male ( $n = 15$ ; 60%). The median age was 54 years (range 7-79 years). Three children (aged seven, 10, and 11 years) were scanned. The reasons for dispatch included road traffic accidents ( $n = 9$ ; 36%), lowered consciousness from other causes ( $n = 9$ ; 36%), falls from heights ( $n = 6$ ; 24%), and blunt trauma to the head ( $n = 1$ ; 4%). The average GCS at the time of the scan was 6.9 (median = 3.0).

In six cases, the scanner was used prior to transporting the patient to the hospital. The other patients ( $n = 19$ ) were scanned during transport to the hospital. The median scan time was four minutes (range 1.5-10 minutes).

The HEMS physicians scored a median ease-of-use of seven (range three to eight). Comments from the users were often related to the difficulty in scanning the dorsal side of the head with the patient in the supine position and to the screen not being visible while scanning. In addition, the Infrascanner produces a sound when the scanning is complete; in a noisy environment, it could be difficult to obtain auditory feedback from the device.

A complete scan, containing all eight scan locations, was conducted in 15 patients (60%). In the other 10 patients, complete scanning of the dorsal occipital site was not possible because of accessibility. These patients were immobilized on a long spine board with head blocks, and because of a suspected cervical spine injury, head manipulation was not conducted.

Data of all the scanned patients ( $n = 25$ ) are presented. The InfraScan results were normal in 10 patients. In 15 patients, the device displayed abnormal results. In one patient, the abnormal scan resulted in an addition to the initial treatment: hypertonic saline (10%) for brain preservation. No patients were transported to a hospital other than the location initially planned because of the InfraScan results.

In the accepting hospitals, a CT scan was performed on all patients, and 11 showed no intracranial bleeding. In 14 patients, the CT scan of the brain showed intracranial bleeding. In four cases, the InfraScan results did not match those of the CT scan. There were three false positive results and one false negative result of the InfraScan.

In the patients younger than 16 years of age ( $n = 3$ ), one false positive result occurred; the other two measurements were adequate (Table 1).

The sensitivity of the Infrascanner in this cohort was 93.3%, with a specificity of 78.6%. When the children were excluded, the sensitivity was the same but the specificity increased to 84.6%.

## Discussion

### *Prehospital Setting*

Early detection and quick, appropriate treatment are essential to prevent secondary damage and to optimize initial treatment in patients with TBI. This approach includes patient triage and transport to a trauma center with adequate neurosurgical intervention options. However, such triage can be challenging, and early detection of intracerebral bleeding may be facilitated with new NIRS devices such as the Infrascanner.

Because this device is relatively new in Europe, the results are presented of an initial cohort of 25 patients scanned on-site in prehospital settings by the physician-staffed HEMS.

### *Feasibility/Usability*

There is a learning curve for using the Infrascanner properly, especially in how to scan in the prehospital setting. Scanning a patient takes time, and this is sometimes not available because of other vital priorities. The median scan time was four minutes (range 1.5-10 minutes). Using this time is defensible when it does not delay vital interventions (such as during transport). Most of the patients (76%) were scanned after initial treatment and stabilization while being transported to the hospital. In these cases, the decision had already been made to transfer the patient to a Level 1 trauma center because of a low GCS, trauma mechanism, or associated injuries. The average GCS at scanning was three. The high incidence of severe TBI in this cohort makes it difficult to draw conclusions regarding the use of the NIRS device as a triage tool in suspected TBI patients. More and less severely injured patients are needed to evaluate this topic and to determine the value of this tool in triage.

The ease-of-use of the Infrascanner was scored; the median score was seven. The observations regarding auditory device feedback and screen/button location will be shared with the manufacturer. If possible, customizing the Infrascanner for noisy prehospital environments may enhance future user satisfaction.

Because many of the patients had possible cervical spine injuries that limited access to the dorsal scan zones, a complete scan of the brain was obtained in 60% of patients. Even with this high number of incomplete scans, a sensitivity for intracerebral hematomas of 93.3% was achieved in the entire cohort.

One patient had a normal InfraScan result but had a small subarachnoid hemorrhage with a volume that was likely to be below the detection level of the Infrascanner. As indicated previously, a minimal amount of blood must be present to be detectable, which was assumed to be more than 3.5 ml and in the first 2.5 cm from the surface of the brain.<sup>10</sup> The larger the area of bleeding is, the more clinically significant it will be. The goal of prehospital scanning is not to detect all minimal hematomas, but to triage patients who eventually need to be transported to a hospital with neurosurgical facilities. Further research is needed to evaluate the detection threshold in relation to clinical intervention.

The obtained sensitivity of 93.3% is slightly higher than that published in the study by Robertson et al (88%). The specificity, at 78.6%, was lower than that previously described in the hospital setting (91%).<sup>10</sup> This result may have been because more extensive bleeding was encountered, well exceeding the detection threshold

Patient No.	Sex	Age	GCS	Complete InfraScan	InfraScan Abnormal	CT Scan Abnormal	InfraScan Matches CT
1	M	59	3	1	0	0	1
2	M	79	3	0	1	1	1
3	F	54	3	0	1	1	1
4	M	38	14	0	0	0	1
5 <sup>a</sup>	V	10	10	1	1	0	0
6 <sup>a</sup>	M	7	3	1	0	0	1
7	M	65	15	0	0	0	1
8	M	37	15	0	0	0	1
9	V	53	3	1	1	1	1
10	V	55	3	0	1	0	0
11	M	66	11	0	0	1	0
12 <sup>a</sup>	M	11	3	0	1	1	1
13	V	34	3	0	0	0	1
14	V	65	3	1	1	1	1
15	V	65	3	1	1	1	1
16	M	73	3	1	0	1	1
17	F	44	15	1	0	0	1
18	M	55	10	1	1	1	1
19	M	74	15	1	1	1	1
20	M	44	3	1	1	0	0
21	M	51	3	1	1	1	1
22	F	67	12	0	1	1	1
23	M	52	3	1	1	1	1
24	M	65	3	1	1	1	1
25	F	44	11	1	0	0	1

Joost © 2017 Prehospital and Disaster Medicine

**Table 1.** Patient Characteristics and Results

Abbreviations: CT, computed tomography; GCS, Glasgow Coma Score.

<sup>a</sup>Data for patients <16 years of age.

mark, which may have increased the number of detections. The relatively large number of false positive results might have been caused by the scanning circumstances. An abnormal NIRS scan result should be verified with a second scan to confirm the reading. In this prehospital study, this was not performed due to the lack of time. Adding a second NIRS scan could further improve the specificity.

In 76% of the cases, patients were scanned while driving at high speeds in ambulances or during helicopter flight. As mentioned before, the back/dorsal side of the patients was frequently not accessible because of the necessity for cervical spine immobilization, which led to a high number of patients with incomplete scans

(40%). This limitation distorted the measurements and may have led to a higher number of false-positive scanning results.

In the authors' opinion, HEMS patients with severe TBI are not the patients who will benefit most from prehospital screening with the Infrascanner. These patients need to go to a center with the highest level of care available, regardless of the scan results. Ambulance services could benefit even more from the NIRS. Further research must focus on the less severely injured patients, who do have an obligatory reason for referral to a Level 1 trauma center, to optimize patient triage and prevent transport delay.

In hospitals, the CT scan is and will remain the gold standard in hematoma screening in TBI patients for the foreseeable future.



However, CT scans use potentially harmful radiation and must therefore be used with caution and only with valid indications, especially in children.<sup>15–17</sup> Thus, alternative algorithms and a step-up approach in cranial screening of TBIs, possibly using NIRS, may be preferable.<sup>12</sup>

Further research regarding cost, time management, triage, patient survival, and quality of life is required to determine the prehospital role of NIRS.

### Limitations

During this research period, many more TBI patients were treated, although the attending HEMS physician had to take the time to perform the investigation. Often, this time was not available or was needed for other vital treatments. If the patient already had an indication to be transported to a Level 1 trauma center, the added value of NIRS scanning for triage was limited in the opinion of the physicians. The decision against performing an InfraScan examination in these cases was not investigated in this study.

In this study, a selection bias is possible because of under-screening of the most severely injured patients, as not all consecutive TBI patients were scanned. In further research, all patients who are eligible for scanning must be included for study and be evaluated.

No long-term patient follow-up is presented. In this pilot study, the aim was to test the InfraScan in a prehospital setting, comparing results with data of the hospital CT scan to assess its practical use and reliability.

### Conclusion

This study obtained promising results using the InfraScan NIRS device in the prehospital screening for intracranial hematomas in TBI patients and found high sensitivity and good specificity. However, further research is necessary to determine the beneficial effects of enhanced prehospital screening on triage, survival, and quality of life in TBI patients.

### Author Contributions

ET and JP designed the study and wrote the draft of the article. JP, BvW, NH, and ET conducted the scans and collected the data used in this article. BvW and NH supported by commenting to article draft and helped with data interpretation. All listed authors made critical contributions to this article regarding its structure and conclusions and provided advice for revising the manuscript. JP submitted the article.

### References

1. Dismuke CE, Walker RJ, Egede LE. Utilization and cost of health services in individuals with traumatic brain injury. *Glob J Health Sci*. 2015;7(6):156-169.
2. Pfeifer R, Tarkin IS, Rocos B, Pape HC. Patterns of mortality and causes of death in polytrauma patients—has anything changed? *Injury*. 2009;40(9):907-911.
3. Brazinova A, Mauritz W, Leitgeb J, et al. Outcomes of patients with severe traumatic brain injury who have Glasgow Coma Scale scores of 3 or 4 and are over 65 years old. *J Neurotrauma*. 2010;27(9):1549-1555.
4. Hoogmartens O, Heselmans A, Van de Velde S, et al. Evidence-based prehospital management of severe traumatic brain injury: a comparative analysis of current clinical practice guidelines. *Prehosp Emerg Care*. 2014;18(2):265-273.
5. Boer C, Franschman G, Loer SA. Prehospital management of severe traumatic brain injury: concepts and ongoing controversies. *Curr Opin Anaesthesiol*. 2012;25(5):556-562.
6. Ghosh A, Elwell C, Smith M. Review article: cerebral near-infrared spectroscopy in adults: a work in progress. *Anesth Analg*. 2012;115(6):1373-1383.
7. Davies DJ, Su Z, Clancy MT, et al. Near-infrared spectroscopy in the monitoring of adult traumatic brain injury: a review. *J Neurotr*. 2015;32(13):933-941.
8. Ding K, Wang H, Wu Y, et al. Rapamycin protects against apoptotic neuronal death and improves neurologic function after traumatic brain injury in mice via modulation of the mTOR-p53-Bax axis. *J Surg Res*. 2015;194(1):239-247.
9. Katzenberger RJ, Chtarbanova S, Rimkus SA, et al. Death following traumatic brain injury in *Drosophila* is associated with intestinal barrier dysfunction. *Elife*. 2015;4.
10. Robertson CS, Zager EL, Narayan RK, et al. Clinical evaluation of a portable near-infrared device for detection of traumatic intracranial hematomas. *J Neurotr*. 2010;27(9):1597-1604.
11. Leon-Carrion J, Dominguez-Roldan JM, Leon-Dominguez U, Murillo-Cabezas F. The Infrascanner, a handheld device for screening in situ for the presence of brain hematomas. *Brain Inj*. 2010;24(10):1193-1201.
12. Salonia R, Bell MJ, Kochanek PM, Berger RP. The utility of near infrared spectroscopy in detecting intracranial hemorrhage in children. *J Neurotr*. 2012;29(6):1047-1053.
13. Bressan S, Daverio M, Martinolli F, et al. The use of handheld near-infrared device (Infrascanner) for detecting intracranial hemorrhages in children with minor head injury. *ChNS*. 2014;30(3):477-484.
14. Braun T, Kunz U, Schulz C, Lieber A, Willy C. [Near-infrared spectroscopy for the detection of traumatic intracranial hemorrhage: feasibility study in a German army field hospital in Afghanistan]. *Unfallchirurg*. 2014.
15. Bajoghli M, Bajoghli F, Tayari N, Rouzbahani R. Children, CT scan, and radiation. *Int J Prev Med*. 2010;1(4):220-222.
16. Hoshiko S, Smith D, Fan C, Jones CR, McNeel SV, Cohen RA. Trends in CT scan rates in children and pregnant women: teaching, private, public, and nonprofit facilities. *Pediatr Radiol*. 2014;44(5):522-528.
17. Blume HK. Children with minor blunt head trauma, a Glasgow Coma Scale score of 14 or 15, and a normal CT scan are at very low risk of traumatic findings on subsequent neuroimaging or of requiring neurosurgical intervention. *Evid Based Med*. 2012;17(6):e11.
18. Tan EC, Peters JH, Hoogerwerf N. Prehospitaal gebruik van een mobiele near-infrared spectroscopie scanner voor diagnostiek bij patiënten met een traumatisch schedelhersensletsel. *Netherlands Mil Med Rev*. 2015;68(2):22-24.

Q2